



protect.ROM telescopic

Fitting instructions for patient

Universal knee splint with flexion / extension limitation and telescopic length adjustment



Do not adjust the dial or loosen the yellow locking buttons on the arms of the brace.



To put the brace on, open the brace fully as shown and lay it behind your leg.



Connect the foam pads around the leg starting at the bottom near the ankle and move up the leg. Make sure the dial is level with your knee.



Tighten the holding straps around the foam pads starting near the ankle again. Check that the dial is still level with your knee.

